

Diaphragm gas meters Pattern evaluation report

Project number : 12200033
Test report number : NMI-12200033-01

Applicant : Apator Metrix S.A.
Piaskowa 3
83-110 Tczew
Poland

Manufacturer : Metrix

Type : UG T

Test specifications : EN 1359:1998/A1:2006 "Diaphragm gas meters"

Testing period : March up to and including April 2012

Issued by : NMI Certin B.V.
Hugo de Grootplein 1
3314 EG DORDRECHT
The Netherlands

Signature :


Ing. H.S. Schouten
Senior Approvals Expert


Ing. J. van der Giessen
Approvals Expert

Date : 2 May 2012

- Tests** : The meters as specified in annex 2 were tested for compliance with the standards as specified on page 1 of this test report. The performed tests are stated in annex 1. If applicable specific test conditions are stated at each test.
- Results** : See annex 1 of this test report. The meter fulfils the class 1,5 requirements of the EN 1359:1998/A1:2006 for all performed tests.
- Based on the compliance with the EN 1359:1998/A1:2006 documents NMI presumes conformity with the Measuring Instrument Directive (MID). The investigation has resulted in a class 1,5 EC type-examination Certificate no. T10382 revision 1.
- Traceability** : The measurements have been executed using standards for which the traceability to (inter)national standards has been demonstrated towards the RvA.
- Uncertainty** : The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, which provides a confidence level of approximately 95%.
The total uncertainty of the measurements of the error of indication is:
- | | at reference conditions: | at other conditions: |
|--------------------------------------|--------------------------|----------------------|
| $Q \geq 400 \text{ dm}^3/\text{h}$: | 0,3% | 0,4% |
| $Q < 400 \text{ dm}^3/\text{h}$: | 0,6% | 0,7% |
- Annexes** : The complete test report consists of the following annexes:
- annex 1 : performed tests
 - annex 2 : characteristics of the tested meters
 - annex 3 : test data
- Remarks** : The examined UG T meter is a revised variant of the previously investigated UG T meter, while using the same measuring system, but with an increased center distance. For the tests which are not performed, as indicated in annex 1, a reference can be made to the previous investigations with the UG T meter, as presented in the test reports NMI-10200983-01 granted by NMI. The results of these tests can be applied also for the UG T with increased center distance.